



Administrative Data

Institution SZX - Institut für Radioonkologie
 Physicist Krauss
 PatientID zz-Inhomogenitätsphantom
 Patient Name zz-Inhomogenitätsphantom, zz-PTW-T40037, zz-Feb-2013
 Comment k(0PTPN6X20AAA)=0.988

Data Set A

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\0PTPN6X20AAA.dcm

Data Set B

Q:\TrueBeam_COMMISSIONING\VeriSoft\2013-02-05 T40037\0PTP6X20d.mcc

Manipulations

Command	Data Set	Parameters	Value
Calibrate Air Density	B	kUser	0.988
Calibrate Air Density	B	kUser	0.989
Calibrate Air Density	B	kUser	0.988

Gamma 2D - Parameters

2.0 mm Distance- To- Agreement
 2.0 % Dose Difference with ref. to Max. dose of measured data set

Statistics

Number of Dose Points	729
Evaluated Dose Points	729 (100.0 %)
Passed	668 (91.6 %)
Failed	61 (8.4 %)
Result	91.6 % (Green)

Settings

Passing criteria	Gamma \leq 1.0
Green	90.0 % to 100.0 %
Yellow	75.0 % to 90.0 %
Red	0.0 % to 75.0 %

